



**STATE OF GEORGIA
DEPARTMENT OF NATURAL RESOURCES
LABORATORY CERTIFICATION**

In accordance with the Georgia Safe Drinking Water Act of 1977 (Sections 12-5-170 through 12-5-193, O.C.G.A.) and the Rules for Safe Drinking Water (Chapter 391-3-5; May 16, 2021)

WAYPOINT ANALYTICAL

*Water Laboratory is hereby granted
Full CERTIFICATION*

*To analyze water samples for compliance with the Revised Total Coliform and Ground Water Rules
using the following methods:*

*Total Coliform: SM 9223B (Colilert), SM9221 D
Escherichia coli: SM 9223B (Colilert), SM9221 F*

*This certification is valid in the State of Georgia and is contingent upon
adherence to the Requirements for Maintaining Certification.*

*April 08, 2025 – April 07, 2026
LABORATORY ID #810*


Laboratory Certification Officer
Drinking Water Compliance Unit
Watershed Protection Branch, EPD


Program Manager
Drinking Water Compliance Unit
Watershed Protection Branch, EPD

CONDITIONS FOR MAINTAINING CERTIFICATION
WAYPOINT ANALYTICAL DRINKING WATER LABORATORY
LABORATORY ID #810

1. Certified laboratories must notify the Laboratory Certification Coordinator by the end of the next business day of any positive test result on samples analyzed for compliance with the Revised Total Coliform Rule or Ground Water Rule.
2. All Total Coliform positive samples must be analyzed specifically for *E. coli* in accordance with the Revised Total Coliform and Ground Water Rules.
3. All staff responsible for analyzing drinking water samples for compliance with the Revised Total Coliform or Ground Water Rules must be certified as Water Laboratory Analyst, Class I Water Operator or Class II Water Operator by the Georgia State Board of Examiners for Certification of Water and Wastewater Treatment Plant Operators and Laboratory Analysts.
4. Certified laboratories are required to participate in Proficiency Testing (PT) studies twice annually (once between January 1 and June 30 and once between July 1 and December 31). Laboratory personnel responsible for analyzing Drinking Water for compliance purposes must individually pass a PT study once a year. PT studies must be completed on both the primary and the secondary method each time. Should a "Not Acceptable" be received for either method, arrangements must be made to participate in a quick turn-around study using that method.
5. Records of coliform bacteria analyses performed for compliance with the Revised Total Coliform Rule or Ground Water Rule must be kept for not less than five (5) years, as prescribed by 391-3-5-.15, of the Georgia Rules for Safe Drinking Water.
6. Certified laboratories must maintain adequate quality control and have these procedures described in their Laboratory Quality Control Manual and Standard Operating Procedures.
7. A downgraded laboratory must note their downgraded status on any report issued during the provisional period.
8. Certified laboratories are required to submit the results of all microbiological compliance samples to the GA EPD.
9. Certified laboratories must notify the Laboratory Certification Coordinator immediately of any change or potential change of laboratory location.
10. Certified laboratories must notify the Laboratory Certification Coordinator within thirty (30) days of major changes in laboratory personnel. The personnel approved to analyze Drinking Water for compliance purposes at this Laboratory are: Kerri Cook, Gayle Shoemake and Dioni Terlaje (Colilert only).



Jeffrey W. Cown, Director

Watershed Protection Branch
2 Martin Luther King, Jr. Drive
Suite 1470A, East Tower
Atlanta, Georgia 30334
404-463-1511

AUG 06 2025

Mr. Jimmy Ferguson
Waypoint Analytical-Carrollton
202 Hamilton Ave.
Bremen, GA 30110

SUBJ: Evaluation of Waypoint Analytical - Carrollton Water Laboratory
Laboratory Identification #810

Dear Mr. Ferguson:

We have received the corrective actions in response to the March 25, 2025, laboratory assessment. The corrective actions have been reviewed and deemed acceptable. Therefore, enclosed you will find your new Certificate. The Certificate has been issued for a one-year period, ending April 07, 2026. You will be contacted by the Laboratory Certification Program prior to that date to schedule your next laboratory assessment.

If you have any questions or need additional information, please feel free to contact me by phone at 470-524-2412 or by e-mail at Jenna.Cook@dnr.ga.gov.

Sincerely,

Jenna Cook
Laboratory Certification Officer
Drinking Water Program
Watershed Protection Branch, GA EPD

Lynne Grubb
Laboratory Certification Officer
Drinking Water Program
Watershed Protection Branch, GA EPD